



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L0917-250414)

This is to certify that

Advantech Co., Ltd.

Advantech QA Laboratory

1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.)

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 0917

Originally Accredited : August 01, 2002

Effective Period : April 28, 2025 to April 27, 2028

Accredited Scope : Testing Field, see described in the Appendix



Scan to verify

Yi-Ling Chen

Yi-Ling Chen
President, Taiwan Accreditation Foundation
April 14, 2025

Accreditation Number : 0917

Laboratory Head : TAI, Mark

■ 19. 99 Electronic and Electric
Electronic and Electric equipment
E003 Cold test
IEC 60068-2-1
MIL-STD-810 Method 502 Procedure I, II
-70 °C to 25 °C

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

E003 Change of temperature test
IEC 60068-2-14 (Nb)
-70 °C to 90 °C
Temperature rate: Guishan District 15 °C/min max., Neihu District 10 °C/min max.

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

E003 Dry heat test
IEC 60068-2-2
MIL-STD-810 Method 501 Procedure I, II
25 °C to 90 °C

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

E003 Mechanical shock test
IEC 60068-2-27
IEC 61373 Category 3
MIL-STD-810 Method 516 Procedure III
Guishan District:
Half- Sine
acceleration: 100 G max. pulse time: 11 ms max.
Square wave
Acceleration 40 G max. Velocity change (256 to 320) ips
Neihu District:
Half- Sine
acceleration: 25 G max. pulse time: 11 ms max.

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

P2, total 3 pages



E003 Damp heat cyclic test
IEC 60068-2-30
MIL-STD-810 Method 507 Procedure II
-70 °C to 90 °C (30 to 95) % R.H.

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

E003 Free fall test
IEC 60068-2-31
MIL-STD-810 Method 516 Procedure IV
High: (30 to 135) cm

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

E003 Vibration test
IEC 60068-2-6
IEC 60068-2-64
IEC 61373 (except category 3))
MIL-STD-810 Method 514 ANNEX C Category 4
MIL-STD-810 Method 514 ANNEX E Category 24
Frequency: 5 Hz to 2000 Hz acceleration: (0 to 10) G

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

E003 Damp heat test
IEC 60068-2-78

Approval Signatory: CHANG, Tim; TSAI, William; TAI, Mark
Testing Site: 1.1F, No.33, Lane 365, Yangguang Street, Neihu District, Taipei City 114, Taiwan (R.O.C.) ; 2. B1, No.27, Wende Road, Guishan District, Taoyuan 333, Taiwan, (R.O.C.)

(Null below)

